



eastwest
medico

DISPOSABLE LATEX GLOVES

TECHNICAL DATA SHEET



PRODUCT INFORMATION

Latex Examination Gloves 24cm. Lightweight, non-sterile, powder-free, medical device Class I, Category III. For single use in the work, safety, hygienic and food sector.



CHARACTERISTICS

Dimensions

Size	Length*	Palm width
XS	Min. 240mm	≤80mm
S	Min. 240mm	80±10mm
M	Min. 240mm	95±10mm
L	Min. 240mm	110±10mm
XL	Min. 240mm	≥110

*in compliance with EN455-2

Thickness - single wall

PALM	FINGER
Min 0,08mm (avg. 0,10mm)	Min 0,11mm (avg. 0,12mm)

LOGISTIC DATA

Gloves in dispenser: 100 gloves

Dispenser dimension: 24 x 12,2 x 6,5 cm

Shelf life: 5 years

Gloves in carton: 10 x 100 gloves

Carton dimension: 24,9 x 34 x 25 cm

Storage: Store in original packaging in a dry and dark place at room temperature. Protect from heat, humidity, bright light and ozone.

DESIGN

Material: Latex

Powder: Powder-free

Shape: Ambidextrous

Cuff: Beaded cuff

Color: Natural (light yellow)

External surface: Palm-textured

Protein content: N/D



Sterilization: Non-sterile

PHYSICAL PROPERTIES

Parameters	Aging before	After aging
Force at break	Minimum 6N	Minimum 6N
Elongation	Minimum 500%	Minimum 400%
AQL	1,5	

QUALITY MANAGEMENT

Product related standards	EN 455-1:2020, EN 455-2:2015, EN 455-3:2023, EN 455-4:2009 EN 374-1:2016, EN 374-1:2016+A1:2018, EN 374-2:2019 EN 374-4:2019, EN 374-5 :2016, EN 21420:2020, EN 16523-1:2015+A12018, ASTM D3578:2019
Quality certificates	ISO 9001:2015, ISO 13485:2016, ISO 14001:2015, ISO 45001:2018
Food safety	Suitable for foodstuffs in accordance to Framework Regulation (EC) No 1935/2004 German Food and Feed Code (LFGB) Recommendation of the German Federal Institute for Risk Assessment (BfR XXI) Category 3 "For short contact with foodstuffs" Not recommended for acidic foodstuffs.
Medical Device	Medical Device, Class 1, in accordance to MDR 2017/745
PPE	Personal protective equipment in accordance to Regulation (EU) 2016/425 CE Category III: For irreversible or mortal risks



BREAKTHROUGH TIMES

Chemical	CAS-no:	Breakthrough times in min. (EN 374-3/EN 16523-1)	Performance Index
Methanol (A)	67-56-1	Not tested	N/A
Acetone (B)	67-64-1	Not tested	N/A
Acetonitrile (C)	75-05-8	Not tested	N/A
Dichloromethane (D)	75-09-6	Not tested	N/A
Carbon disulphide (E)	71-15-0	Not tested	N/A
Toluene (F)	108-88-3	Not tested	N/A
Diethylamine (G)	109-89-7	Not tested	N/A
Tetrahydrofuran (H)	109-99-9	Not tested	N/A
Ethyl Acetate (I)	141-78-6	Not tested	N/A
n-Heptane (J)	142-82-5	Not tested	N/A
Sodium Hydroxide 40% (K)	1310-73-2	>480	Level 6
Sulfuric acid 96% (L)	7664-93-9	20	Level 1
Nitric acid 65% (M)	7697-37-2	25	Level 1
Acetic acid 99% (N)	64-19-7	Not tested	N/A
Ammonia 25% (O)	1336-21-6	Not tested	N/A
Hydrogen peroxide 30% (P)	7722-84-1	>480	Level 6
Hydrofluoric acid 40% (S)	7664-39-3	Not tested	N/A
Formaldehyde 37% (T)	50-00-0	Between 361 to 480	Level 5

PLEASE NOTICE

Glove length is appropriate to the use where the risk is minimal to the wrist area.

USAGE

The product needs to be checked for the correct size and suitability with regard to its intended use. Before use check for damage and do not use damaged or unsuitable gloves. The gloves are intended for single use only and shall be disposed on a regular basis.

CAUTION

The product may cause allergic reactions on sensitive people. Particular caution must be exercised if the user has any known or suspected hypersensitivity or allergy towards natural rubber latex.

The information above are to be considered as typical values in the standard version based on the manufacturer's information. Subject to modification. The gloves shall only be used for its intended purpose.

Revision 7. Date of revision: 4/06/2024.

All errors and print errors reserved.

ewmedico.com

PRODUCTION

Manufactured in Thailand by Sri Trang Gloves (Thailand) PCL for Prosenso ApS/Eastwest Medico ApS., Denmark.

TECHNICAL DATA SHEET ISSUED BY

Eastwest Medico Aps
Gotfred Rodes Vej 2
2920 Charlottenlund
Denmark

REORDERING

Item no.	Size	Number	Boxes per carton
EWLTX55MST1-XS	XS	100 pieces/box	10
EWLTX55MST1-S	S	100 pieces/box	10
EWLTX55MST1-M	M	100 pieces/box	10
EWLTX55MST1-L	L	100 pieces/box	10
EWLTX55MST1-XL	XL	100 pieces/box	10

Prosenso ApS own the name and logo trademarks in this datasheet, and adheres strictly to the international ISO 13485 standard for manufacturing of Medical Devices.